



# California Medical Device Recall Information



## Recall Name

**Playtex Recalls AC/DC Power Adapters Used With the Playtex®  
Nurser Deluxe Double Electric Breast Pump  
Due to Potential for Electric Shock**

Recall Date	Product Description	Recalling Firm	Recall Reason
Update: 09/09/14  Initial: 03/18/14	AC/DC Power Adapters  Used with the Playtex® Nurser Deluxe Double Electric Breast Pump	<b>Energizer Personal Care - Playtex Manufacturing, Inc.</b> Shelton, CT	<i>The casing on some adapters may become loose and separate.</i>  <b><i>Potential for electric shock.</i></b>
Recall Class	Product Identification	Distribution	Affected Dates
I	Serial Numbers:  • <b>P12315-XXXX, through P13205-XXXX</b>  [or by Adapter Production Codes:  <b>1238 through 1324.</b> ]  <a href="#">Product Photos</a>	<b>CA</b> , nationwide	Manufactured  November 2012 through July 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm414564.htm>